

Certificate of Analysis

Product Information:

Orion Number	203020
Description	SILICA STANDARD 200 PPB, 5 L
Lot Number	HR1
Expiration Date	9/04
Method of Analysis	UV Spectrophotometry
Acceptance Range	200 ppb +/- 10 ppb Sodium Metasilicate Nonahydrate (as SiO ₂)

Actual Analysis:

<u>Analysis</u>	<u>Value</u>	<u>Unit of Measure</u>
1	203.9	ppb Silica

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
NOT APPLICABLE	NA	NA

Certificate Date: 09/23/03

Approved By:

Edward Henning

Quality Control Supervisor