

Certificate of Analysis

Product Information:

Orion Number	203020
Description	SILICA STANDARD 200 PPB, 5 L
Lot Number	HO1
Expiration Date	12/04
Method of Analysis	UV Spectrophotometer
Acceptance Range	200 ppb +/- 10 ppb Sodium Metasilicate Nonahydrate (as SiO ₂)

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	208.2	ppb Silica

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
NOT APPLICABLE	NA	NA

Certificate Date: 12/12/03

Approved By:



Quality Control Supervisor