

# Certificate of Analysis

## Product Information:

**Orion Number** 910107  
**Description** # PH BUFFER 7, 475ML  
**Lot Number** FU2  
**Expiration Date** 06/30/03  
**Method of Analysis** Potentiometric

## Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	7.000	pH

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Disodium Hydrogen Phosphate	186 II	f
Potassium Dihydrogen Phosphate	186 I	f
Sodium Bicarbonate	191	b
Sodium Carbonate	192	b

**Certificate Date:** 07/10/01

**Approved By:**



**Quality Control Supervisor**

Environmental Instruments  
Water Analysis

166 Cummings Center

Beverly, MA  
01915

(978) 232-6015 Dom Fax  
(978) 232-6031 Int'l Fax

www.thermo.com  
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

**Thermo**  
ELECTRON CORPORATION

245967-001 Rev. B