

Certificate of Analysis

Product Information:

Orion Number 203020
Description SILICA STANDARD 200 PPB, 5 L
Lot Number DO1
Expiration Date 12/31/00
Method of Analysis UV Spectrophotometry

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	198.6	ppb Silica

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
NOT APPLICABLE		NA

Certificate Date: 02/23/00

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

245967-001 Rev. B

Thermo
ELECTRON CORPORATION