

Certificate of Analysis

Product Information:

Orion Number 910725
Description PH 7 BUFFER IN 25 PACK.
Lot Number DS3
Expiration Date 08/31/01
Method of Analysis Potentiometric

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	7.003	pH

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Dihydrogen Phosphate	186 I	f
Disodium Hydrogen Phosphate	186 II	f
Sodium Carbonate	192	b
Sodium Bicarbonate	191	b

Certificate Date: 12/08/99

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

Thermo
ELECTRON CORPORATION

245967-001 Rev. B