

# Certificate of Analysis

## Product Information:

**Orion Number** 181140-1  
**Description**  
**Lot Number** CY1  
**Expiration Date** 02/28/00  
**Method of Analysis** Flame Photometry

## Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	19.2	ppm Sodium Chloride as Sodium

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

**Certificate Date:** 10/21/99

**Approved By:**



**Quality Control Supervisor**

Environmental Instruments  
Water Analysis

166 Cummings Center

Beverly, MA  
01915

(978) 232-6015 Dom Fax  
(978) 232-6031 Int'l Fax

www.thermo.com  
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

245967-001 Rev. B

**Thermo**  
ELECTRON CORPORATION