

Certificate of Analysis

Product Information:

Orion Number	150072
Description	REF. ELEC. IFS FOR 100058 5 X 60ML BOTTLES
Lot Number	LU1
Expiration Date	2 years after receipt of product
Method of Analysis	Potentiometric
Acceptance Range	0.100 +/- 0.005 M Cesium Bromide

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	.099	Moles/Liter Cesium Bromide

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

Certificate Date: 06/15/07

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION