

Certificate of Analysis

Product Information:

| | |
|---------------------------|--|
| Orion Number | 177011 |
| Description | !REAGENT 900ML W/54 G REAGENT POWDER 30 DAY SUPPLY |
| Lot Number | IW1 |
| Expiration Date | 2 Years after Receipt of Product |
| Method of Analysis | Potentiometric |
| Acceptance Range | 9.03 M +/- 0.20 M Acetic Acid |

Actual Analysis:

| <u>Analysis</u> | <u>Results</u> | <u>Unit of Measure</u> |
|-----------------|----------------|-------------------------|
| 1 | 9.03 | Moles/Liter Acetic Acid |

Traceability:

| <u>N.I.S.T. Chemical</u> | <u>SRM #</u> | <u>Sublot</u> |
|------------------------------|--------------|---------------|
| Potassium Hydrogen Phthalate | 185 | h |

Certificate Date: 04/29/04

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. B

Thermo
ELECTRON CORPORATION