

Certificate of Analysis

Product Information:

Orion Number	182040
Description	CALCIUM STANDARDS KIT Standard 1
Lot Number	IQ1
Expiration Date	10/06
Method of Analysis	Potentiometric
Acceptance Range	1.91 mM+/- 0.01 mM Calcium Chloride

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	1.90	milliMoles/Liter Calcium Chloride

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Calcium Carbonate	915a	NA

Certificate Date: 10/27/04

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. B

Thermo
ELECTRON CORPORATION