

Certificate of Analysis

Product Information:

Orion Number	931911
Description	POTASSIUM ISA, 475 ML
Lot Number	IO1
Expiration Date	2 Years after Receipt of Product
Method of Analysis	Potentiometric
Acceptance Range	5.0 +/- 0.1 M NaCL

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	5.0	Moles/Liter Sodium Chloride

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

Certificate Date: 12/13/04

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. B

Thermo
ELECTRON CORPORATION