

# Certificate of Analysis

## Product Information:

<b>Orion Number</b>	900010
<b>Description</b>	REF ELEC FILL SOLN 5 X 60 ML
<b>Lot Number</b>	JT1
<b>Expiration Date</b>	2 Years after Receipt of Product
<b>Method of Analysis</b>	Potentiometric
<b>Acceptance Range</b>	2.00 M +/- 0.10 M Ammonium Chloride

## Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	1.98	Moles/Liter Ammonium Chloride

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

**Certificate Date:** 07/15/05

**Approved By:**



**Quality Control Supervisor**

**Environmental Instruments**  
Water Analysis Instruments

166 Cummings Center

Beverly, MA  
01915

(978) 232-6000  
(978) 232-6015 fax

[www.thermo.com/water](http://www.thermo.com/water)  
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. C

**Thermo**  
ELECTRON CORPORATION