

Certificate of Analysis

Product Information:

| | |
|---------------------------|--|
| Orion Number | 150072 |
| Description | REF. ELEC. IFS FOR 100058 5 X 60ML BOTTLES |
| Lot Number | JO1 |
| Expiration Date | 2 Years after Receipt of Product |
| Method of Analysis | Potentiometric |
| Acceptance Range | 0.100 +/- 0.005 M Cesium Bromide |

Actual Analysis:


| <u>Analysis</u> | <u>Results</u> | <u>Unit of Measure</u> |
|-----------------|----------------|----------------------------|
| 1 | 0.099 | Moles/Liter Cesium Bromide |

Traceability:

| <u>N.I.S.T. Chemical</u> | <u>SRM #</u> | <u>Sublot</u> |
|--------------------------|--------------|---------------|
| Sodium Chloride | 2201 | NA |

Certificate Date: 01/02/06

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

Thermo
ELECTRON CORPORATION

244721-001 Rev. C