

Certificate of Analysis

Product Information:

Orion Number	203020
Description	SILICA STANDARD 200 PPB, 5 L
Lot Number	KV1
Expiration Date	05/07
Method of Analysis	UV Spectrophotometer
Acceptance Range	200.0 ppb+/-10.0 ppb SiO ₂

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	200	ppb Silica

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
NOT APPLICABLE	NA	NA

Certificate Date: 05/18/06

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION