

Certificate of Analysis

Product Information:

Orion Number	181140
Description	SODIUM STANDARDS KIT Std. 1
Lot Number	KX1
Expiration Date	3/08
Method of Analysis	Flame Photometer
Acceptance Range	19.1 ppm +/- 0.5 ppm Sodium Chloride as Sodium

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	19.20	ppm Sodium

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

Certificate Date: 03/28/06

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION