

Certificate of Analysis

Product Information:

| | |
|---------------------------|----------------------------------|
| Orion Number | 181811 |
| Description | !REAGENT W/TUBING 3XL BTLS |
| Lot Number | KQ1 |
| Expiration Date | 2 Years after Receipt of Product |
| Method of Analysis | Potentiometric |
| Acceptance Range | 13.1 M +/- 0.6 M Acetic Acid |

Actual Analysis:

| <u>Analysis</u> | <u>Results</u> | <u>Unit of Measure</u> |
|-----------------|----------------|-------------------------|
| 1 | 13.7 | Moles/Liter Acetic Acid |

Traceability:

| <u>N.I.S.T. Chemical</u> | <u>SRM #</u> | <u>Sublot</u> |
|------------------------------|--------------|---------------|
| Potassium Hydrogen Phthalate | 185H | NA |

Certificate Date: 10/31/06

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION