

# Certificate of Analysis

## Product Information:

<b>Orion Number</b>	182040-1
<b>Description</b>	CALCIUM STANDARDS KIT
<b>Lot Number</b>	NNV1
<b>Expiration Date</b>	05/2011
<b>Method of Analysis</b>	Potentiometric
<b>Acceptance Range</b>	1.91+/-0.01mM CaCl

## Actual Analysis:


<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	1.92	Moles/Liter Calcium Chloride

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Calcium Carbonate	915b	NA

**Certificate Date:** 07/08/09

**Approved By:**



**Quality Representative**

### Water Analysis Instruments

#### North America

166 Cummings Center  
Beverly, MA 01915 USA  
Toll Free: 1-800-225-1480  
Tel: 1-978-232-6000  
Dom. Fax: 1-978-232-6015  
Int'l Fax: 978-232-6031

#### Europe

P.O. Box 254, 3860 AG Nijkerk  
Wallerstraat 125K, 3862 BN  
Nijkerk, Netherlands  
Tel: (31) 033-2463887  
Fax: (31) 033-2460832

#### Asia Pacific

Blk 55, Ayer Rajah Crescent  
#04-16/24, Singapore 139949  
Tel: 65-6778-6876  
Fax: 65-6773-0836

[www.thermo.com/water](http://www.thermo.com/water)

**Thermo**  
SCIENTIFIC