

# Certificate of Analysis

## Product Information:

**Orion Number** 181811  
**Description** !REAGENT W/TUBING 3XL BTLS  
**Lot Number** NS1  
**Expiration Date** N/A; Recommendation usage 2 years from date of receipt  
**Method of Analysis** Potentiometric  
**Acceptance Range** 13.1 M +/- 0.6 M Acetic Acid

## Actual Analysis:


<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	13.0	Moles/Liter Acetic Acid

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Hydrogen Phthalate	185h	NA

**Certificate Date:** 09/11/09

**Approved By:**



**Quality Representative**

### Water Analysis Instruments

#### North America

166 Cummings Center  
Beverly, MA 01915 USA  
Toll Free: 1-800-225-1480  
Tel: 1-978-232-6000  
Dom. Fax: 1-978-232-6015  
Int'l Fax: 978-232-6031

#### Europe

P.O. Box 254, 3860 AG Nijkerk  
Wallerstraat 125K, 3862 BN  
Nijkerk, Netherlands  
Tel: (31) 033-2463887  
Fax: (31) 033-2460832

#### Asia Pacific

Blk 55, Ayer Rajah Crescent  
#04-16/24, Singapore 139949  
Tel: 65-6778-6876  
Fax: 65-6773-0836

[www.thermo.com/water](http://www.thermo.com/water)

**Thermo**  
SCIENTIFIC