

Certificate of Analysis

Product Information:

| | |
|---------------------------|---------------------------------------|
| Orion Number | 203020 |
| Description | SILICA STANDARD 200 PPB, 5 L |
| Lot Number | OZ1 |
| Expiration Date | 01/2011 |
| Method of Analysis | Spectrophotometer |
| Acceptance Range | 200.0 ppb+/-10.0 ppb SiO ₂ |

Actual Analysis:


| <u>Analysis</u> | <u>Results</u> | <u>Unit of Measure</u> |
|-----------------|----------------|------------------------|
| 1 | 203.8 | ppb Silica |

Traceability:

| <u>N.I.S.T. Chemical</u> | <u>SRM #</u> | <u>Sublot</u> |
|--------------------------|--------------|---------------|
| NOT APPLICABLE | N/A | NA |

Certificate Date: 02/01/10

Approved By:



Quality Representative

Water Analysis Instruments

North America

166 Cummings Center
Beverly, MA 01915 USA
Toll Free: 1-800-225-1480
Tel: 1-978-232-6000
Dom. Fax: 1-978-232-6015
Int'l Fax: 978-232-6031

Europe

P.O. Box 254, 3860 AG Nijkerk
Wallerstraat 125K, 3862 BN
Nijkerk, Netherlands
Tel: (31) 033-2463887
Fax: (31) 033-2460832

Asia Pacific

Blk 55, Ayer Rajah Crescent
#04-16/24, Singapore 139949
Tel: 65-6778-6876
Fax: 65-6773-0836

www.thermo.com/water

Thermo
SCIENTIFIC