

Certificate of Analysis

Product Information:

Orion Number	110902
Description	STANDARD FLUORIDE, 1000 PPM
Lot Number	MT1
Expiration Date	07/2010
Method of Analysis	Potentiometric
Acceptance Range	0.0526 M +/- 0.0005 M NaF

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	0.0528	Moles/Liter Sodium Fluoride
2	1003	ppm Fluoride

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Fluoride	2203	NA

Certificate Date: 07/14/08

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION