



Material Safety Data Sheet

Product: CEDIA[®] Mycophenolic Acid Calibrator Kit

Catalog #: 100277

Description:

Mycophenolic Acid Calibrator Low Liquid

Mycophenolic Acid Calibrator High Liquid

Chemical Hazard/Hazardous Ingredients:

N/A

Fire, Explosion and Reactivity Data:

Flashpoint: No data.

Extinguishing media: Use extinguishing media appropriate for surrounding fire.

Unusual fire and explosion hazard: None.

Reactivity: Stable.

Hazardous polymerization: Will not occur.

Incompatibility: Avoid copper or lead plumbing. May form potentially explosive metal azides.

Health Hazards:

HARMFUL IF SWALLOWED. The toxicological properties of this product have not been determined.

First Aid:

Ingestion - wash out mouth with water, provided the person is conscious. Call physician.

Inhalation - remove to fresh air. If breathing becomes difficult, give oxygen and call physician.

Skin contact - remove contaminated clothing and flush the contacted area with water and wash thoroughly with soap and water. If irritation persists, call physician.

Eye contact - flush with copious amounts of water for 15 minutes. Assure adequate flushing by separating the eyelids with fingers. Call physician.

Continued on next page

Microgenics Corporation

46360 Fremont Blvd., Fremont, CA 94538 USA ☐ Tel: (510) 979-5000 ☐ Fax: (510) 979-5498
Technical/Customer Service (800) 232-3342



Material Safety Data Sheet (cont.)

Spill and Leak Procedure:

Wear appropriate protective equipment. Sweep up or absorb material and place in a closed container. Ventilate and wash spill area.

If material is disposed of using plumbing, always flush with a large volume of water to prevent azide build-up.

Disposal should be made in accordance with existing disposal practices employed for waste at your institution. Observe all federal, state and local laws.

Special Protection Information:

Provide adequate general mechanical and local exhaust ventilation. Protect eyes and skin with safety glasses and gloves. Avoid contact with eyes and skin. Do not breathe solution vapor.

Special Storage Precautions:

Avoid high temperatures (200 °F).

Do not store with acids.

Doc: 100277MSDS-0

Date: 2006 10

Microgenics Corporation

46360 Fremont Blvd., Fremont, CA 94538 USA □ Tel: (510) 979-5000 □ Fax: (510) 979-5498
Technical/Customer Service (800) 232-3342