



Material Safety Data Sheet

Product: CEDIA[®] THC Plus Calibrators & Controls

Catalog #: 10010032, 10010033, 10010034, 10010035, 10010036, 10010037, 10010038

Description:

CEDIA THC Plus 25 Calibrator	Liquid
CEDIA THC Plus 50 Calibrator	Liquid
CEDIA THC Plus 100 Calibrator	Liquid
CEDIA THC Plus 200 Calibrator	Liquid
CEDIA THC Plus 25 Control Set Low	Liquid
CEDIA THC Plus 25 Control Set High	Liquid
CEDIA THC Plus 50 Control Set Low	Liquid
CEDIA THC Plus 50 Control Set High	Liquid
CEDIA THC Plus 100 Control Set Low	Liquid
CEDIA THC Plus 100 Control Set High	Liquid

Chemical Hazard/hazardous Ingredients: Human urine.

Fire, Explosion and Reactivity Data:

Flashpoint - Not determined.

Extinguishing media - Use extinguishing media appropriate for surrounding fire.

Unusual fire and explosion hazard - None

Reactivity - Stable. Hazardous polymerization will not occur.

Incompatibility - Avoid copper or lead plumbing. May form potentially explosive metal azides.

Health Hazards: Although the human urine from which this product is derived was found negative for Hepatitis B surface antigen and antibody to HTLV III/LAV by approved test method, the product should be HANDLED AS IF IT IS POTENTIALLY INFECTIOUS.

HARMFUL IF SWALLOWED. The toxicological properties of this product has not been determined.

Continued on next page

Microgenics Corporation

46360 Fremont Blvd., Fremont, CA 94538 USA □ Tel: (510) 979-5000 □ Fax: (510) 979-5498
Technical/Customer Service (800) 232-3342

Material Safety Data Sheet (cont.)

First Aid:

Ingestion - wash out mouth with water, provided the person is conscious. Call physician.

Inhalation - remove to fresh air. If breathing becomes difficult, give oxygen and call physician.

Skin contact - remove contaminated clothing and flush the contacted area with water and wash thoroughly with soap and water. If irritation persists call physician.

Eye contact - flush with copious amounts of water for 15 minutes. Assure adequate flushing by separating the eyelids with fingers. Call physician.

Spill and Leak Procedure: Wear appropriate protective equipment. Sweep up or absorb material and place in a closed container. Ventilate and wash spill area. If material is disposed of using plumbing always flush with a large volume of water to prevent azide build-up. Disposal should be made in accordance with existing disposal practices employed for waste at your institution. Observe all federal, state and local laws.

Special Protection Information: Provide adequate general mechanical and local exhaust ventilation. Protect eyes and skin with safety glasses and gloves. Avoid contact with eyes and skin. Do not breathe solution vapor.

Special Storage Precautions:

Avoid high temperatures (200 °F)

Do not store with acids

Doc: 10010032MSDS

Date: 2006 12

Microgenics Corporation

46360 Fremont Blvd., Fremont, CA 94538 USA ☐ Tel: (510) 979-5000 ☐ Fax: (510) 979-5498
Technical/Customer Service (800) 232-3342