

Certificate of Analysis

Product Information:

Orion Number	011006
Description	COND/TDS STD 12.9MS/CM 5x60ML
Lot Number	MQ1
Expiration Date	10/2010
Method of Analysis	Conductivity
Acceptance Range	12.9 mS/cm +/-0.07 mS/cm@25.0 deg.

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	12.87	mS/cm at 25 deg. C

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Chloride	2202	NA

Certificate Date: 10/10/08

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION