

# Certificate of Analysis

## Product Information:

<b>Orion Number</b>	150072
<b>Description</b>	REF. ELEC. IFS FOR 100058 5 X 60ML BOTTLES
<b>Lot Number</b>	MQ1
<b>Expiration Date</b>	2 Years after Receipt of Product
<b>Method of Analysis</b>	Potentiometric
<b>Acceptance Range</b>	0.100 +/- 0.005 M Cesium Bromide

## Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	.099	Moles/Liter Cesium Bromide

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

**Certificate Date:** 10/31/08

**Approved By:**



**Quality Representative**

**Environmental Instruments**  
Water Analysis Instruments

166 Cummings Center

Beverly, MA  
01915

(978) 232-6000  
(978) 232-6015 fax

www.thermo.com/water  
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

**Thermo**  
ELECTRON CORPORATION