

LCQ DECA XP Exhaust Requirements

It is your responsibility as the user to provide an adequate exhaust system. Much of what is introduced into the LCQ Deca XP MS detector will eventually be exhausted from the rotary-vane pumps, along with the small amount of oil vapor that these pumps characteristically emit. Therefore, the pumps should be connected to a fume exhaust system.

Note. An efficient fume exhaust system is required for the proper operation of your rotary-vane pumps. Most API applications will contribute to the accumulation of solvents in the rotary-vane pumps. These solvents must be purged from the mechanical pump oil periodically by opening the ballast valves located on the top of the pumps. When the ballast valves are opened a large volume volatile solvent waste might enter the fume exhaust system. Therefore, your fume exhaust system must be able to accommodate the periodic purging of the solvents. The frequency of the purging is dependent on the throughput of your system.

The rotary-vane pumps have two functions: (1) provide forepressure for the turbomolecular pump, (2) provide a vacuum for the capillary skimmer of the API source.

Each rotary-vane pump requires a 25 mm (1 in.) exhaust port.