

## EC DECLARATION OF CONFORMITY

We,

**Thermo Electron Oy, Clinical Chemistry & Automation Systems**

declare that the Chemistry Analyzers with

Trade name: **Konelab Arena 30**  
Code: **984161**  
Type: **969**

are in conformity with the provisions of the following EC directives, including the latest amendments, and with national legislation implementing these directives:

73/23/EEC: Low Voltage Directive  
89/336/EEC: EMC Directive

and that the following harmonized standards have been applied:

EN 61010-1  
EN 61326-1  
EN 61000-6-3  
EN 61000-6-1

This Declaration is valid for all batches which are placed on the market by ourselves on or after October 20<sup>th</sup>, 2004 and which bear the CE marking. This analyzer is **not** intended to be used for IVD purposes.

Vantaa, March 21, 2006

Thermo Electron Oy



Anne-Maria Peitsalo  
Quality Manager  
Quality, Regulatory and Compliance  
Clinical Chemistry & Automation Systems