

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : D-Glucose + D-Fructose + Sucrose

Catalogue no. : 984317

Lot no. : G588

Package : R1: 3x16ml R2: 3x4.5ml R3: 3x4.5ml R4: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2013-01


### Quality Control

#### Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control 1	73.36	71.94 – 87.92
Control 2	9.46	8.99 – 10.99

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 10 / 8 2011

  
\_\_\_\_\_  
Hannele Laine  
Quality Assurance Specialist  
Quality, Regulatory and Compliance  
Clinical Diagnostics Finland

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : D-Glucose + D-Fructose + Sucrose

Catalogue no. : 984301

Lot no. : G587

Package : R1: 3x45ml R2: 3x13ml R3: 3x13ml R4: 3x13ml

Storage Conditions : 2...8 °C

Expiry date : 2013-01

### Quality Control

#### Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control 1	76.56	71.94 – 87.92
Control 2	9.71	8.99 – 10.99

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 5 / 8 2011



Marja Keurulainen  
Quality Chemist  
Quality, Regulatory and Compliance  
Clinical Diagnostics Finland

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : D-Glucose + D-Fructose + Sucrose

Catalogue no. : 984317

Lot no. : G153

Package : R1: 3x16ml R2: 3x4.5ml R3: 3x4.5ml R4: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2012-06

### Quality Control

#### Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control 1	73.67	71.87 – 87.84
Control 2	9.47	8.98 – 10.98

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 25 / 2 2011



Hannele Laine  
Quality Assurance Specialist  
Quality, Regulatory and Compliance  
Clinical Diagnostics Finland

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : D-Glucose + D-Fructose + Sucrose

Catalogue no. : 984317

Lot no. : F622

Package : R1: 3x16ml R2: 3x4.5ml R3: 3x4.5ml R4: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2011-12

### Quality Control

#### Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control 1	76.26	71.87 – 87.84
Control 2	9.70	8.98 – 10.98

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 31 / 8 2010



Hannele Laine  
Quality Assurance Specialist  
Quality, Regulatory and Compliance  
Clinical Diagnostics Finland

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : D-Glucose + D-Fructose + Sucrose

Catalogue no. : 984317

Lot no. : 3217

Package : R1: 3x14ml R2: 3x4.8ml R3: 3x4.8ml R4: 3x4.8ml

Storage Conditions : 2...8 °C

Expiry date : 2011-12


### Quality Control

#### Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control 1	76.61	71.49 – 87.37
Control 2	9.63	8.94 – 10.92

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 11 / 3 2010

  
Hannele Laine  
Quality Assurance Specialist  
Quality, Regulatory and Compliance  
Clinical Diagnostics Finland

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : D-Glucose + D-Fructose + Sucrose

Catalogue no. : 984317

Lot no. : 2562

Package : R1: 3x14ml R2: 3x4.8ml R3: 3x4.8ml R4: 3x4.8ml

Storage Conditions : 2...8 °C

Expiry date : 2011-11

### Quality Control

#### Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control 1	75.87	71.62 – 87.54
Control 2	9.55	8.96 – 10.95

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 15 / 12 2009



Marja Keurulainen  
Quality Chemist  
Quality, Regulatory and Compliance  
Clinical Diagnostics Finland