

ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : Ammonia

Catalogue no. : 984320

Lot no. : G595

Package : R1: 3x16ml R2: 3x4.5ml R3: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2012-11

Quality Control

Performance test:

<u>Control</u>	<u>Result (mg/l)</u>	<u>Target range (mg/l)</u>
Control 1	61.0	54 – 66
Control 2	20.4	18 – 22

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 9 / 8 2011



Hannele Laine
Quality Assurance Specialist
Quality, Regulatory and Compliance
Clinical Diagnostics Finland

ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : Ammonia

Catalogue no. : 984320

Lot no. : G326

Package : R1: 3x16ml R2: 3x4.5ml R3: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2012-07

Quality Control

Performance test:

<u>Control</u>	<u>Result (mg/l)</u>	<u>Target range (mg/l)</u>
Control 1	60.3	54 – 66
Control 2	21.0	18 – 22

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 21 / 4 2011



Hannele Laine
Quality Assurance Specialist
Quality, Regulatory and Compliance
Clinical Diagnostics Finland

ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : Ammonia

Catalogue no. : 984320

Lot no. : G137

Package : R1: 3x16ml R2: 3x4.5ml R3: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2012-04

Quality Control

Performance test:

<u>Control</u>	<u>Result (mg/l)</u>	<u>Target range (mg/l)</u>
Control 1	61.2	54 – 66
Control 2	20.3	18 – 22

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 21 / 4 2011



Hannele Laine
Quality Assurance Specialist
Quality, Regulatory and Compliance
Clinical Diagnostics Finland

ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy
Ratatie 2, FIN-01620 Vantaa, Finland

Product Name : Ammonia

Catalogue no. : 984320

Lot no. : G101

Package : R1: 3x16ml R2: 3x4.5ml R3: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2012-04

Quality Control

Performance test:

<u>Control</u>	<u>Result (mg/l)</u>	<u>Target range (mg/l)</u>
Control 1	60.71	54 – 66
Control 2	21.00	18 – 22

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 21 / 4 2011



Hannele Laine
Quality Assurance Specialist
Quality, Regulatory and Compliance
Clinical Diagnostics Finland