

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name: Ethanol

Catalogue no.: 984300

Lot no.: G944

Package R1: 3x16ml R2: 3x4.5ml

Storage Conditions: 2...8 °C

Expiry date: 2013-03


### Quality Control

#### Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control 1 (0.5g/l)	0.55	0.4 - 0.6
Control 2 (1.0g/l)	1.07	0.9 - 1.1
Control 3 (2.0g/l)	2.11	1.8 - 2.2
Control 4 (3.0g/l)	3.14	2.7 - 3.3

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 12 / 12 2011

  
\_\_\_\_\_  
Hannele Laine  
Quality assurance Specialist  
Quality, Regulatory and Compliance  
Clinical Diagnostics Finland

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : Ethanol

Catalogue no. : 984300

Lot no. : G322

Package : R1: 3x16ml R2: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2012-06


### Quality Control

#### Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control 1 (0.5g/l)	0.51	0.4 - 0.6
Control 2 (1.0g/l)	1.03	0.9 - 1.1
Control 3 (2.0g/l)	2.03	1.8 - 2.2
Control 4 (3.0g/l)	3.10	2.7 - 3.3

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 20 / 4 2011

  
\_\_\_\_\_  
Hannele Laine  
Quality assurance Specialist  
Quality, Regulatory and Compliance  
Clinical Diagnostics Finland