

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratatie 2, FIN-01620 Vantaa, Finland

Product Name: D-Glucose + D-Fructose

Catalogue no.: 984314

Lot no.: G911

Package : R1: 3x16ml R2: 3x4.5ml R3: 3x4.5ml

Storage Conditions: 2...8 °C

Expiry date: 2013-04

### Quality Control

#### Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control High	59.6	54.0 – 66.0
Control Low	6.0	5.4 – 6.6

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 1 / 12 2011



Hannele Laine  
Quality Assurance Specialist  
Quality, Regulatory and Compliance  
Clinical Diagnostics Finland

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : D-Glucose + D-Fructose

Catalogue no. : 984313

Lot no. : G636

Package : R1: 3x45ml R2: 3x13ml R3: 3x13ml

Storage Conditions : 2...8 °C

Expiry date : 2013-04


### Quality Control

#### Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control 1	59.04	53.78 – 65.74
Control 2	5.96	5.38 – 6.58

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 29/8 2011

  
Hannele Laine  
Quality Assurance Specialist  
Quality, Regulatory and Compliance  
Clinical Diagnostics Finland

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : D-Glucose + D-Fructose

Catalogue no. : 984314

Lot no. : G250

Package : R1: 3x16ml R2: 3x4.5ml R3: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2012-12

### Quality Control

#### Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control 1	60.19	53.90 – 65.88
Control 2	6.20	5.39 – 6.59

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 23 / 0 2011



Marja Keurulainen  
Quality Chemist  
Quality, Regulatory and Compliance  
Clinical Diagnostics Finland

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : D-Glucose + D-Fructose

Catalogue no. : 984314

Lot no. : G011

Package : R1: 3x16ml R2: 3x4.5ml R3: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2012-06


### Quality Control

#### Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control 1	61.19	53.99 – 65.99
Control 2	6.13	5.40 – 6.60

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 11 / 1 2011

  
Hannele Laine  
Quality Assurance Specialist  
Quality, Regulatory and Compliance  
Clinical Diagnostics Finland

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : D-Glucose + D-Fructose

Catalogue no. : 984314

Lot no. : F757

Package : R1: 3x16ml R2: 3x4.5ml R3: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2012-07

### Quality Control

#### Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control 1	59.93	53.87 – 65.85
Control 2	6.05	5.39 – 6.59

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 18 / 10 2010



Hannele Laine  
Quality Assurance Specialist  
Quality, Regulatory and Compliance  
Clinical Diagnostics Finland