

ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : Glycerol

Catalogue no. : 984316

Lot no. : G691

Package : R1: 3x16ml R2: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2013-01


Quality Control

Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control High	9.91	9.09 – 11.11
Control Low	1,03	0.91 – 1.11

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 6/9 2011



Hannele Laine
Quality Assurance Specialist
Quality, Regulatory and Compliance
Clinical Diagnostics Finland

ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy
Ratatie 2, FIN-01620 Vantaa, Finland

Product Name : Glycerol

Catalogue no. : 984316

Lot no. : G151

Package : R1: 3x16ml R2: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2012-10


Quality Control

Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control 1	9.69	9.01 – 11.01
Control 2	0.96	0.90 – 1.10

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 25 / 2 2011



Hannele Laine
Quality Assurance Specialist
Quality, Regulatory and Compliance
Clinical Diagnostics Finland