

ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : L-Lactic Acid

Catalogue no. : 984308

Lot no. : G634

Package : R1: 3x16ml R2: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2013-03


Quality Control

Performance test:

<u>Control</u>	<u>Result (mg/l)</u>	<u>Target range (mg/l)</u>
Control 1	136	132 - 162
Control 2	27	26 - 32

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 29 / 8 2011


Hannele Laine
Quality Assurance Specialist
Quality, Regulatory and Compliance
Clinical Diagnostics Finland

ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : L-Lactic Acid

Catalogue no. : 984308

Lot no. : G215

Package : R1: 3x16ml R2: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2012-06

Quality Control

Performance test:

<u>Control</u>	<u>Result (mg/l)</u>	<u>Target range (mg/l)</u>
Control 1	138	134 - 164
Control 2	31	27 - 33

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 11 / 3 2011



Hannele Laine
Quality Assurance Specialist
Quality, Regulatory and Compliance
Clinical Diagnostics Finland

ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : L-Lactic Acid

Catalogue no. : 984308

Lot no. : F798

Package : R1: 3x16ml R2: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2012-07


Quality Control

Performance test:

<u>Control</u>	<u>Result (mg/l)</u>	<u>Target range (mg/l)</u>
Control 1	142	132 - 162
Control 2	31	26 - 32

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa / / 2010



Hannele Laine
Quality Assurance Specialist
Quality, Regulatory and Compliance
Clinical Diagnostics Finland

ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : L-Lactic Acid

Catalogue no. : 984308

Lot no. : 3453

Package : R1: 3x18ml R2: 3x4.8ml

Storage Conditions : 2...8 °C

Expiry date : 2012-03

Quality Control

Performance test:

<u>Control</u>	<u>Result (mg/l)</u>	<u>Target range (mg/l)</u>
Control 1	136	132 - 162
Control 2	28	26 - 32

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 11 / 6 2010


Hannele Laine
Quality Assurance Specialist
Quality, Regulatory and Compliance
Clinical Diagnostics Finland

ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : L-Lactic Acid

Catalogue no. : 984308

Lot no. : 3077

Package : R1: 3x18ml R2: 3x4.8ml

Storage Conditions : 2...8 °C

Expiry date : 2012-01


Quality Control

Performance test:

<u>Control</u>	<u>Result (mg/l)</u>	<u>Target range (mg/l)</u>
Control 1	136	133 - 163
Control 2	27	27 - 33

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 11 / 2 2010


Hannele Laine
Quality Assurance Specialist
Quality, Regulatory and Compliance
Clinical Diagnostics Finland