

ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : Sucrose (Total Glucose)

Catalogue no. : 984312

Lot no. : G776

Package : R1: 3x16ml R2: 3x4.5ml R3: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2012-06


Quality Control

Performance test:

Control	Result (g/l)	Target range (g/l)
Control 1	24.22	22.55 - 27.55
Control 2	2.37	2.01 - 3.01

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 18 / 10 2011


Päivi Junell
Quality Chemist
Quality, Regulatory and Compliance
Clinical Diagnostics Finland

ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : Sucrose (Total Glucose)

Catalogue no. : 984312

Lot no. : G040

Package : R1: 3x16ml R2: 3x4.5ml R3: 3x4.5ml

Storage Conditions : 2...8 °C

Expiry date : 2012-06


Quality Control

Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control 1	23.38	22.55 - 27.56
Control 2	2.36	2.01 - 3.01

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 26 / 1 2011


Hannele Laine
Quality Assurance Specialist
Quality, Regulatory and Compliance
Clinical Diagnostics Finland

ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : Sucrose (Total Glucose)

Catalogue no. : 984312

Lot no. : 3454

Package : R1: 3x14ml R2: 3x4.8ml R3: 3x4.8ml

Storage Conditions : 2...8 °C

Expiry date : 2012-01


Quality Control

Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control 1	24.88	22.39 - 27.37
Control 2	2.49	1.99 - 2.99

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 24/5 2010



Hannele Laine
Quality Assurance Specialist
Quality, Regulatory and Compliance
Clinical Diagnostics Finland

ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : Sucrose (Total Glucose)

Catalogue no. : 984312

Lot no. : 3078

Package : R1: 3x14ml R2: 3x4.8ml R3: 3x4.8ml

Storage Conditions : 2...8 °C

Expiry date : 2012-01

Quality Control

Performance test:

<u>Control</u>	<u>Result (g/l)</u>	<u>Target range (g/l)</u>
Control 1	24.47	22.38 - 27.36
Control 2	2.49	1.99 - 2.99

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 11 / 2 2010



Hannele Laine
Quality Assurance Specialist
Quality, Regulatory and Compliance
Clinical Diagnostics Finland